- 11 Emery RW, Van Nooten GJ, Tesar PJ. The initial experience with the ATS Medical
- mechanical cardiac valve prosthesis. Ann Thorac Surg 2003;75:444–52.
 Bernal JM, Martin-Duran R, Rabasa J, et al. The Carbomedics "Top Hat" supra-annular prosthesis. Ann Thorac Surg 1999;67:1299–303.
 Gelsomino S, Morocutti G, Da Col P, et al. Preliminary experience with the St
- Jude Medical Regent mechanical heart valve in the aortic position: early in vivo hemodynamic results. Ann Thorac Surg 2002;73:1830-6.

 14 Bach D, Goldbach M, Sakwa MP, et al. Hemodynamics and early performance
- of the St Jude Medical Regent aortic valve prosthesis. J Heart Valve Dis 2001:10:436-42.
- Wiseth R, Haaverstad R, Vitale N, et al. Prosthetic valve haemodynamics assessed by the left ventricular outflow tract utilization index: a randomized study of the Carbomedics reduced versus Metronic Hall valve. J Heart Valve Dis
- 16 Wiseth R, Levang OW, Sande E, et al. Hemodynamic evaluation by Doppler echocardiography of small (≤21 mm) prostheses and bioprostheses in the aortic valve position. Am J Cardiol 1992;**70**:240–6.
- **Rothbart RM**, Castriz JL, Harding LV, *et al*. Determination of aortic valve area by two-dimensional and Doppler echocardiography in patients with normal stenotic bioprosthetic valves. *J Am Coll Cardiol* 1990;15:817–24.

- 18 Vitale N, Cappabianca G, Visicchio G, et al. Midterm evaluation of the Sorin Bicarbon heart valve prosthesis: single-center experience. Ann Thorac Surg 2004·**77**·527-31
- Zeller JL, Kratz JM, Crumbley AJ, et al. Long-term experience with the St. Jude Medical valve prosthesis. Ann Thorac Surg 1999;**68**:1210–18.

 20 **Bernal JM**, Rabasa JM, Gutierrez-Garcia F, et al. The Carbomedics valve:
- experience with 1,049 implants. Ann Thorac Surg 1998;65:137-43.
- 21 Eichinger WB, Wagner I, Bleiziffer S, et al. Occasional single beat regurgitation observed with the new Medtronic Advantage bileaflet heart valve. J Thorac Cardiovasc Surg 2006;131:730-1.
- Baumgartner H, Khan SS, DeRobertis M, et al. Doppler assessment of prosthetic valve orifice area. An in vitro study. Circulation 1992;85:2275–83.
 Mascherbauer J, Schima H, Maurer G, et al. Doppler assement of mechanical
- aortic valve prostheses: effect of valve design and size of the aorta. J Heart Valve Dis 2004;13:823-30.
- 24 Shandas R, Kwon J, Valdes-Cruz L. A method for determining the reference effective flow areas for mechanical heart valve prostheses. In vitro validation study. Circulation 2000:101:1953-9
- 25 Rowlands DB Glover DR Ireland MA et al. Assessment of left ventricular mass and its response to antihypertensive treatment. Lancet 1982;2:467-70.


IMAGES IN CARDIOLOGY.....

64 multidetector CT in patent foramen ovale

47-year-old woman was referred for a coronary CT angiography (64-slice scanner, Toshiba Aquilion, Tustin, California, USA). Retrospective ECG-gated datasets were acquired (collimation, 64×0.5 mm; table feed per rotation, 7.2 mm; rotation, 400 ms; tube voltage, 120 kV; and tube current, 500 mA). Using a segmented image reconstruction algorithm, a temporal resolution of <100 ms was achieved. Before scan, the heart rate was lowered to 65 bpm by 50 mg oral metoprolol. Contrast enhancement was achieved with 65 ml of Iohexol (Omnipaque 350 mg/ml, Amersham Health, Cork, Ireland) injected at 5 ml/s, followed by an injection of 50 ml of saline at 5 ml/s. Scanning initiation was triggered automatically at the threshold of 180 Hounsfield Units in the descending aorta. Using this technique, the right heart was washed out of contrast by saline chaser. This provides enough contrast to demonstrate small shunt across the interatrial septum. Presented images show a small patent foramen ovale with the jet of contrast moving from the left to the right atrium.

> F Saremi, S F Attai, J Narula fsaremi@uci.edu

doi: 10.1136/hrt.2006.097782



Long- and short-axis images through the interatrial septum show a small patent foramen ovale (arrows) with the jet of contrast from the left atrium (LA) to the right atrium (RA). The right heart is less dense due to washout of contrast by saline chaser. Cardiac images are obtained on the basis of relative timing intervals in steps of 10% of the R-R interval in cardiac cycle. The centre of reconstruction is placed at the relative point as shown in the diagram (top right). On the left side, high-resolution long- and short-axis views are reformatted from diastolic datasets centred at 70% of the R-R interval, with a slice thickness of 0.5 mm and increments of 0.3 mm. On the right, short-axis cine frames (2 mm thickness) from collected data at 10%, 30%, 60% and 90% intervals are shown